



**Mahidol University**  
Faculty of Medical Technology

# Preliminary Report

The External Quality Assessment Scheme in Clinical Immunology (EQAI)

## Hepatitis B Virus Serology

Trial: 4 / 2024

Issue No. 1 Issue Date: 21 November 2024

Sample ID	Preliminary consensus values from participants					
	HBsAg	AntiHBs		AntiHBc	HBeAg	AntiHBe
	Qualitative	Qualitative	Quantitative (mIU/ML) Median, nIQR	Qualitative	Qualitative	Qualitative
SE02-24-4-1	Negative	Negative	2.14, 0.91	Negative	Negative	Negative
SE02-24-4-2	Negative	NS	78.01, 8.49	NS	Negative	NS
SE02-24-4-3	Positive	Negative	2, 0.5	Positive	Positive	Negative
SE02-24-4-4	Negative	NS	120.15, 18.88	NS	Negative	NS

Qualitative data based on all methods  
AntiHBs quantitative data based on E-CLIA method  
NS= Non-consensus

**Provided by**

EQAI : Hepatitis B Virus Serology  
Faculty of Medical Technology, Mahidol University  
999 Salaya, Phutthamonthon, Nakhon Pathom 73170  
Phone: 02 441 4376 Ext. 2529  
Mobile: 080 982 5570  
Email: eqai.mahidol@gmail.com

**Approved by:**

Assoc. Prof. Chaniya Leepiyasakulchai, Ph.D.  
Scheme Coordinator