



Mahidol University
Faculty of Medical Technology

Preliminary Report

The External Quality Assessment Scheme in Clinical Immunology (EQAI)

Hepatitis B Virus Serology

Trial: 2 / 2025

Issue No. 1 Issue Date: 19 June 2025

Sample ID	Preliminary consensus values from participants					
	HBsAg	AntiHBs		AntiHBc	HBeAg	AntiHBe
	Qualitative	Qualitative	Quantitative (mIU/ML) X_{PT} , σ_{PT}	Qualitative	Qualitative	Qualitative
SE02-25-2-1	Positive	Negative	2.33, 0.73	Positive	Negative	Positive
SE02-25-2-2	Negative	Negative	4.56, 1.84	Negative	Negative	Negative
SE02-25-2-3	Negative	NS	27.57, 3.56	Negative	Negative	Negative
SE02-25-2-4	Positive	Negative	2.59, 1.46	Positive	Negative	Positive

Qualitative data based on all methods
AntiHBs quantitative data based on E-CLIA method
NS= Non-consensus

Provided by

EQAI : Hepatitis B Virus Serology
Faculty of Medical Technology, Mahidol University
999 Salaya, Phutthamonthon, Nakhon Pathom 73170
Phone: 02 441 4376 Ext. 2529
Mobile: 080 982 5570
Email: eqaihbv.mtmu@gmail.com

Approved by:

Assoc. Prof. Chaniya Leepiyasakulchai, Ph.D.
Scheme Coordinator